



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/763,975	01/22/2004	Brian J. Cox	388700-058B	7891
37374 7590 12/27/2010 INSKEEP INTELLECTUAL PROPERTY GROUP, INC 2281 W. 190TH STREET SUITE 200 TORRANCE, CA 90504				
EXAMINER SEVERSON, RYAN J				
ART UNIT 3731		PAPER NUMBER		
NOTIFICATION DATE 12/27/2010		DELIVERY MODE ELECTRONIC		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

inskeepstaff@inskeeplaw.com

Office Action Summary

Application No.

10/763,975

Applicant(s)

COX, BRIAN J.

Examiner

RYAN J. SEVERSON

Art Unit

3731

Period for Reply -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 09 November 2010.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 23-43 is/are pending in the application.
- 4a) Of the above claim(s) 29-39, 42 and 43 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 23-28, 40 and 41 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftperson's Patent Drawing Review (PTO-945)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Response to Amendment

1. Examiner notes here that in response to the Board's new grounds of rejection, applicant has amended the independent claims. This is an indication that prosecution has been voluntarily reopened by the applicant. The rejection set forth by the Board has been treated as a non-final rejection, in which applicant was given an opportunity to respond. Applicant has done so and successfully overcome the rejection the Board set forth. However, the amendment necessitated an updated search in which new prior art has been discovered. Because the amendments necessitated the new grounds of rejection set forth below, this action is being made final. See MPEP 1214.01 and 37 CFR 41.50(b).

Claim Rejections - 35 USC § 103

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. **Claims 23-28, 40 and 41 are rejected under 35 U.S.C. 103(a) as being unpatentable over Deem et al. (6,231,597) in view of Boock (6,309,367).** Deem et al. disclose a device that is a support structure (see figure 4) sized for placement at an aneurysm (see figure 11B). The support structure has a bridge portion or occlusion region (15, see figure 1) that spans the neck of the aneurysm (see figure 11A). The support structure has an open configuration (see figure 4).

4. However, the embodiment of Deem et al. described above does not disclose the support structure is non-tubular. Attention is drawn to figure 13 of Deem et al., which shows a support structure that does not form a complete loop (see column 8, lines 8-17), which would be beneficial because the support structure does not obstruct as much of the lumen, thereby reducing the resistance to blood flow through the area in which the support structure is placed. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to make the embodiment in figures 1 and 11A of Deem et al. with the support structure that does not encompass the entire circumference of the lumen in the manner taught in figure 13 so as not to obstruct as much of the lumen, thereby reducing the resistance to blood flow through the area in which the support structure is placed.
5. Further regarding claims 23 and 40, Deem et al. disclose substantially identical embodiments in figures 12 and 13, wherein the only difference is the end portions extend around the entire circumference of the lumen in figure 12 and do not in figure 13. This is further evidence that the support structure can perform equally well configured in either fashion and there would be no disadvantage to modifying the embodiment of figures 1 and 11A to have the end portions extend only about a portion of the circumference of the lumen.
6. Deem et al. fail to disclose a volumetrically expandable reactive material on the bridge portion. Attention is drawn to Boock, who teaches a swellable (volumetrically expandable) material (94, see figure 13) on the central portion (analogous to the bridge portion of Deem et al.) of an aneurysm treatment device to provide a seal at the

aneurysm neck that prevents fluid from passing into the aneurysm (see column 4, lines 15-21 of Boock). Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to have included a volumetrically expandable reactive material on the bridge portion of Deem et al. as suggested by Boock to seal the neck of the aneurysm.

7. Regarding claims 24, 25 and 41, the arced configuration of Deem et al. is curved and coiled (see figure 1) and conforms to the lumen it is placed in (see figure 11A).

8. Regarding claim 26, the support structure of Deem et al. includes a sinusoidal body portion (elements 14 form a sinusoidal pattern, see figure 1).

9. Regarding claim 27, the sinusoidal pattern of Deem et al. is only disposed in the bridge portion, which lies between the opposing ends of the support structure.

10. Regarding claim 28, the bridge portion of the combination of Deem et al. and Boock includes the reactive material (as described above in paragraph 6).

Response to Arguments

11. Applicant's arguments with respect to claims 23 and 40 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

12. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP

§ 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

13. A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

14. Any inquiry concerning this communication or earlier communications from the examiner should be directed to RYAN J. SEVERSON whose telephone number is (571)272-3142. The examiner can normally be reached on Monday - Friday 8:30-5:00.

15. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anh Tuan Nguyen can be reached on (571) 272-4963. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

16. Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Ryan J Severson/
Examiner, Art Unit 3731
12/16/10

/(Jackie) Tan-Uyen T. Ho/
Supervisory Patent Examiner, Art Unit 3773